




গণ বিশ্ববিদ্যালয়
সাভার, ঢাকা।

দরপত্র আহ্বান

তারিখ: ০২/০৪/২০২৪ খ্রিঃ

গণ বিশ্ববিদ্যালয়ের ফার্মেসী বিভাগের জন্য সংযুক্ত তালিকায় উল্লেখিত যন্ত্রপাতি সমূহ (Lot -B) ক্রয়ের নিমিত্তে প্রকৃত ও বৈধ সরবরাহকারী প্রতিষ্ঠানের নিকট থেকে দরপত্র আহ্বান করা যাচ্ছে। সংশ্লিষ্ট কাজে অভিজ্ঞ প্রতিষ্ঠানসমূহকে সংযুক্ত তালিকা অনুযায়ী দরপত্র সীলমোহরকৃত খামে অথবা লেটারহেড প্যাডে সীল স্বাক্ষরসহ এবং ট্রেড লাইসেন্স এর কপিসহ গণ বিশ্ববিদ্যালয়ের স্টোর অফিসে রক্ষিত টেন্ডার বক্রে আগামী ১৭/০৪/২০২৪ ইং তারিখে দুপুর ১:০০ ঘটিকার মধ্যে জমা প্রদানের জন্যে অনুরোধ করা যাচ্ছে। উল্লেখিত তারিখে দুপুর ২:০০ ঘটিকায় উপস্থিত দরদাতাদের (যদি কেউ থাকেন) সামনে সকল দরপত্র উন্মুক্ত করা হবে।

উল্লেখ্য যে, কর্তৃপক্ষ কোন কারণ দর্শানো ব্যতিরেকে যে কোন সময় সকল দরপত্র বাতিল বা গ্রহণ করার ক্ষমতা সংরক্ষণ করে।

 ২.৫.২৪

(অধ্যাপক ড. মোঃ নুরুল ইসলাম)

সভাপতি

টেন্ডার ও কোটেশন কমিটি

সদয় জ্ঞাতার্থে-

১. উপাচার্য, গণ বিশ্ববিদ্যালয়।
২. কোষাধ্যক্ষ, গণ বিশ্ববিদ্যালয়।
৩. রেজিস্ট্রার, গণ বিশ্ববিদ্যালয়।
৪. পরিচালক, (অর্থ ও হিসাব বিভাগ), গণ বিশ্ববিদ্যালয়।
৫. বিভাগীয় ক্রয় কমিটি, ফার্মেসী বিভাগ, গণ বিশ্ববিদ্যালয়।
৬. সিনিয়র সহকারী রেজিস্ট্রার, গণ বিশ্ববিদ্যালয়।
৭. আইটি বিভাগ (ওয়েবসাইটে আপলোডের জন্য), গণ বিশ্ববিদ্যালয়।
৮. সকল নোটিশ বোর্ড, গণ বিশ্ববিদ্যালয়।
৯. স্টোর কর্মকর্তা, গণ বিশ্ববিদ্যালয়।



ফার্মেসী বিভাগের যন্ত্রপাতির চাহিদা তালিকাঃ

Lot- B

ক্রমিক নং	যন্ত্রপাতির নাম	স্পেসিফিকেশন	প্রস্তুতকারক ও দেশ	পরিমাণ/ সংখ্যা
1.	Hotplate Magnetic Stirrer	<ul style="list-style-type: none">Model: HPT-380PHeating temperature range: Room temp. $\pm 5^{\circ}\text{C}$-380°COver heat protection: 420°CTemperature display accuracy: $\pm 0.1^{\circ}\text{C}$External temperature sensor: PT1000 (accuracy $\pm 0.5^{\circ}\text{C}$)	USA/Taiwan/ Korea	2 Pc
2.	Forced Air Drying Oven	<ul style="list-style-type: none">Model: DOF- HAS70ACapacity: 70LTemperature range: RT+10-250 $^{\circ}\text{C}$Safety device: Leakage protector/Over- temp. alarm	USA/Taiwan/ Korea	2 Pc
3.	4 Digit balance with high efficiency	<ul style="list-style-type: none">Readability: 0.0001gMax. Capacity: 100gExternal CalibrationTareFunction/CountingOverload Alarm	Japan/Germany/ Taiwan /USA/Korea	1 Pc
4.	3-digit Weighing Balance	<ul style="list-style-type: none">Capacity Maximum Weight: 300 gmAccuracy: 0.001gAuto zero trackingLow battery indicationBack-light displayHigh accuracy load cellLarge LCD displayAuto power off to save battery lifeUpdate speed: 7.5 Hz	Japan/Germany/ Taiwan /USA/Korea	4 Pc

[Handwritten signature]

5.	Top loading rough balance	Weighing range: 1gm - 600 gm	Japan/Germany/ Taiwan /USA/Korea	4 Pc
6.	Water bath	<ul style="list-style-type: none"> • 1 chamber with 4 openings, 2 rows • Volume: 14.6L • Temp. Resolution Ratio: 0.1°C 	USA/ Germany	4 Pc
7.	Heating Mantle with Regulator Control	<ul style="list-style-type: none"> • Flask Size: 2 Litre • Watt 500 • Max. Temperature 450°C • Dimensions : 29 x 12 cm • Weight: 3.2 kg 	Singapore/ Korea	2 pc
8.	Heating Mantle with Regulator Control	<ul style="list-style-type: none"> • Flask Size 3 Litre • Max. Temperature 450°C • Watt: 500 	Singapore/ Korea	2 pc
9.	Diaphragm Vacuum Pump	<ul style="list-style-type: none"> • Pumping Speed: 20 L/Min • Ultimate Vacuum: 200 mbar 	USA/Germany/Taiwan	1 Pc
10.	Cone blender SS	<ul style="list-style-type: none"> • SS Pharma grade • Capacity: 2-5 kg 	India/China/ Taiwan/Korea	1 Pc
11.	Sieving set with electric vibrator	Mesh size: 60 and 80	India/China/ Taiwan/Bangladesh	1 Pc
12.	Mass Mixture (Horizontal)	<ul style="list-style-type: none"> • SS Pharma grade • Capacity: 2-5 kg 	India/China/ Taiwan/Bangladesh	1 Pc
13.	Non-Invasive Tail Cuff Plethysmography	<ul style="list-style-type: none"> • Analog output • Typical Pulse amplitude 100mV to 1 V • Pressure amplitude adjustable 1-5V at 300 mmHg • -120/240V CE 	India/China/ Taiwan/Italy	1 Pc
14.	Auto Photometer	<ul style="list-style-type: none"> • Model No. FP640N/ FP6410 • Sensitivity: K:0.01ppm; Na:0.01ppm • Measuring range: K:0-100ppm; Na:0-160ppm • Linearity: K:0.195ppm(0.39-3.12) ppm Na:0.69 ppm(1.15-9.2)ppm 	India/China/ Taiwan/Italy	1 Pc

		<ul style="list-style-type: none"> • Repeatability:1% CV for 20 consecutive samples 		
15.	Analgesiometer	<ul style="list-style-type: none"> • Power Requirements : 115 or 230V, 50/60Hz, 15W max. • Start / Stop : by pedal switch • Force Range 37215 : 0 to 250, 500, 750 grams 37216 : 0 to 125, 250, 375 grams 	India/China/ Taiwan/Italy	1 pc

Dr. M. N. Islam 2.4.24

(অধ্যাপক ড. মোঃ নুরুল ইসলাম)

সভাপতি

টেভার ও কোটেশন কমিটি